

## Certificate of Analysis

---

---

Product Name: **96-well white plate, sterile**  
Lot Number: **B19103NY**

Storage Conditions: **Not Applicable**  
Expiration Date: **12/01/22**


**FOR RESEARCH USE ONLY**  
**CAUTION: Not intended for human or animal therapeutic use.**

### Quality Control

Parameter	Specification	Results
Raw Materials	high quality/income controlled	Pass
DNase / RNase / human DNA free	negative by quant. PCR amplification	Negative
Endotoxin free	≤ 0.03EU/ml	Pass

---

Manufactured and tested for Preferred Cell Systems™ by Supplier

  
\_\_\_\_\_  
Paul Bork, QA Manager  
Date *3/21/2020*